

CERTIFICATE OF ANALYSIS

Theileria equi Antibody Test Kit, cELISA

Catalog No.:	274-2
Lot:	P230427-001
Expiration:	12 June 2025
Note: For this COA to be valid the expiration date must be equivalent to that of the kit box label.	

QC Testing	Date of QC	POS % I	Mean POS OD	Mean NEG OD
Potency QC	12 June 2023	74.0	0.252	0.971

This product has met both VMRD, Inc. and USDA requirements for serial release.

QC Manager,



David Hyatt

13 June 2023