

CERTIFICATE OF ANALYSIS

Canine Coombs

Positive Control Canine anti-SRBC

Catalog No.:	372-2
Lot:	P230417-002
Expiration:	08 June 2025
Subagglutination Dose:	1/12

Quality Control Method: Washed sheep red blood cells (SRBC) were sensitized with the reagent at the subagglutinating dose and tested as a positive control with the Coombs Reagent (catalog number 392-2 or 392-5).

Specific Reaction: The subagglutinating dose for the SRBCs used for QC was determined to be 1/12. At this dilution there was a 1+ positive on the sensitized SRBCs at 1/4 dilution of the Coombs Reagent.

Other Comments: The subagglutinating dose for this antiserum was determined to be 1/12. This dose may vary under your laboratory conditions, and especially with your source of sheep red blood cells. Therefore, we recommend that you titer this antiserum with your own SRBCs before using as your positive control. Please refer to Section A of the Coombs Positive Control procedure.

QC Manager:



David Hyatt

Technical Data Sheet Version: 1

FOR *IN VITRO* LABORATORY USE ONLY.

15 June 2023