

CERTIFICATE OF ANALYSIS

# Bluetongue Virus Antibody Test Kit, cELISA

Catalog No.:	5010.20-2
Lot:	P210601-002
Expiration:	11 April 2023
Note: For this COA to be valid the expiration date must be equivalent to that of the kit box label.	

QC Testing	Date of QC	POS % I	Mean POS OD	Mean NEG OD
Potency QC	11 October 2021	78.3	0.213	0.980

This product has met both VMRD, Inc. and USDA requirements for serial release.

QC Manager,



David Hyatt

14 October 2021