

CERTIFICATE OF ANALYSIS

Feline Leukemia Virus (FeLV)

Secondary FITC Conjugate

Catalog No.:	AB2-FELV
Lot:	P170714-001
Expiration:	26 July 2021

Quality Control Method: Indirect FA using FeLV 2-well control slide (catalog no. SLD-FAC-FELV), positive blood smear (catalog no. SLD-BSP-FELV), negative blood smear (catalog no. SLD-BSN-FELV), and primary anti-FeLV antibody (catalog no. AB1-FELV).

Specific Reaction: 2-4+ signal on the positive well, negative on the negative well with trace background on both wells. 1-3+ signal on the positive blood smear, negative on the negative blood smear.

Other Comments: NA

QC Manager: Chris Cay



QC Release Date: 02 August 2017

Pattern of Fluorescence: On the 2 well slide there should be Individual cell cytoplasmic fluorescence with occasional small syncytia.

FA Cross-Reaction Screening: The raw material has also been tested satisfactorily by VMRD's IFA systems for *Bartonella henselae*, Feline Calicivirus (FCV), Feline Leukemia Virus (FeLV), Feline Herpesvirus (FHV-1/FVR), Feline Immunodeficiency Virus (FIV), Feline Infectious Peritonitis type 1 and 2 (FIP-1 and 2), Feline Panleukopenia Virus (FPLV), Reovirus (REO), and Rabies Recombinant Nucleoprotein (rNP), *Toxoplasma gondii*, Vesicular Stomatitis Virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: Version #1

FOR *IN VITRO* LABORATORY USE ONLY.

H:\Master Documents\COA (VMRD)\Conjugate\FelV\FelV Secondary Conjugate\AB2-FELV_P170714-001-210726.docx

05 July 2018