

CERTIFICATE OF ANALYSIS

# Feline Leukemia Virus (FeLV)


Secondary FITC Conjugate

Catalog No.:	AB2-FELV
Lot:	P200820-001
Expiration:	23 September 2024

**Quality Control Method:** Indirect FA using FeLV 2-well control slide (catalog no. SLD-FAC-FELV), positive blood smear (catalog no. SLD-BSP-FELV), three negative blood smears (catalog no. SLD-BSN-FELV), and primary anti-FeLV antibody (catalog no. AB1-FELV).

**Specific Reaction:** 1–3+ signal, trace–1 background on the positive well and negative, trace–1 background on the negative well. 1–3+ signal, no background on the positive blood smear and negative, no background on the negative blood smear.

**Other Comments:** NA

**QC Manager:** 

David Hyatt

**Pattern Of Fluorescence:** On the 2 well slide there should be Individual cell cytoplasmic fluorescence with occasional small syncytia.

**FA Cross-Reaction Screening:** The raw material has also been tested satisfactorily by VMRD's IFA systems for *Bartonella henselae*, Feline Calicivirus (FCV), Feline Herpesvirus (FHV-1/FVR), Feline Infectious Peritonitis type 1 and 2 (FIP-1 and 2), Feline Panleukopenia Virus (FPLV), Feline Immunodeficiency Virus (FIV), Rabies Recombinant Nucleoprotein (rNP), Reovirus (REO), *Toxoplasma gondii*, Vesicular Stomatitis Virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: 1

FOR *IN VITRO* LABORATORY USE ONLY.

12 October 2020