

CERTIFICATE OF ANALYSIS

Bovine Adenovirus Type 1 (BAV-1)

FITC Conjugate

Catalog No.:	CJ-F-BAV1-10ML
Lot:	P180502-003
Expiration:	17 May 2022

Quality Control Method: Direct FA using BAV-1 in house 12-well slide

Specific Reaction: 1-3+ signal with no background on positive cells and negative signal with no background on negative cells.

Other Comments: NA

QC Manager: Chris Cay

QC Release Date: 21 May 2018



Pattern Of Fluorescence: Individual cells with diffuse cytoplasmic and membrane fluorescence and strong nuclear fluorescence.

FA Cross-Reaction Screening: The raw material has also been screened by FA and does not react with Bovine Adenovirus Type 3, and 5 (BAV-3, and 5), Bovine Coronavirus (BCV), Bovine Leukemia Virus (BLV), Bovine Parvovirus (BPV), Bovine Respiratory Syncytial Virus (BRSV), Bluetongue Virus (BTV), Bovine Viral Diarrhea Virus (BVDV), Bovine Herpesvirus type 1 (BHV-1/IBR), Parainfluenza Virus Type 3 (PI-3), Reovirus (REO), Rabies Recombinant Nucleoprotein (rNP), *Toxoplasma gondii*, Vesicular Stomatitis Virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: Version # 2

FOR *IN VITRO* LABORATORY USE ONLY.

23 May 2018