

CERTIFICATE OF ANALYSIS

## Bovine Coronavirus (BCV)

FITC Conjugate

Catalog No.:	CJ-F-BCV-10ML
Lot:	P211203-001
Expiration:	20 December 2025

Quality Control Method: Direct FA using BCV 2-well slide (SLD-FAC-BCV).

Specific Reaction: 2–4+ signal on the positive well and negative on the negative well with trace background.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Granular cytoplasmic and nuclear fluorescence mainly on single cells but also some small syncytia.

FA Cross-Reaction Screening: The raw material has also been screened by FA and does not react with Bovine Adenovirus Type 1, 3, and 5 (BAV-1, 3, and 5), Bovine Herpesvirus type 1 (BHV-1/IBR), Bovine Leukemia Virus (BLV), Bovine Parvovirus (BPV), Bovine Respiratory Syncytial Virus (BRSV), Bluetongue Virus (BTV), Bovine Viral Diarrhea Virus (BVDV), *Neospora caninum*, Parainfluenza Virus Type 3 (PI-3), Rabies Recombinant Nucleoprotein (rNP), Reovirus (REO), *Toxoplasma gondii*, Vesicular Stomatitis Virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: 1

FOR IN VITRO LABORATORY USE ONLY.

21 December 2021