

CERTIFICATE OF ANALYSIS

Bovine Parvovirus (BPV)

FITC Conjugate

Catalog No.:	CJ-F-BPV-10ML
Lot:	P180831-001
Expiration:	17 October 2022

Quality Control Method: Direct FA using BPV 2-well slide (catalog no. SLD-FAC-BPV).

Specific Reaction: 2 – 4+ signal on the positive well with trace background, negative on the negative well with trace background.

Other Comments: NA

QC Manager: Chris Cay



QC Release Date: 17 October 2018

Pattern Of Fluorescence: Ground glass and/or granular cytoplasmic fluorescence with occasional bright and diffuse nuclear fluorescence.

FA Cross-Reaction Screening: The raw material has also been screened by FA and does not react with Bovine Adenovirus Type 1, 3, and 5 (BAV-1, 3, and 5), Bovine Coronavirus (BCV), Bovine Leukemia Virus (BLV), Bovine Respiratory Syncytial Virus (BRSV), Bluetongue Virus (BTV), Bovine Viral Diarrhea Virus (BVDV), Bovine Herpesvirus type 1 (BHV-1/IBR), Parainfluenza Virus Type 3 (PI-3), Reovirus (REO), Rabies Recombinant Nucleoprotein (rNP), Vesicular Stomatitis Virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: Version # 1

FOR *IN VITRO* LABORATORY USE ONLY.

18 October 2018