

CERTIFICATE OF ANALYSIS

Bovine Parvovirus (BPV)


FITC Conjugate

Catalog No.:	CJ-F-BPV-10ML
Lot:	P201113-001
Expiration:	11 December 2024

Quality Control Method: Direct FA using BPV 2-well slide (catalog no. SLD-FAC-BPV).

Specific Reaction: 3–4+ signal on the positive well with no background. Negative on the negative well, trace background.

Other Comments: NA

QC Manager: 

David Hyatt

Pattern Of Fluorescence: Ground glass and/or granular cytoplasmic fluorescence with occasional bright and diffuse nuclear fluorescence.

FA Cross-Reaction Screening: The raw material has also been screened by FA and does not react with Bovine Adenovirus Type 1, 3, and 5 (BAV-1, 3, and 5), Bovine Coronavirus (BCV), Bovine Herpesvirus type 1 (BHV-1/IBR), Bovine Leukemia Virus (BLV), Bovine Respiratory Syncytial Virus (BRSV), Bluetongue Virus (BTV), Bovine Viral Diarrhea Virus (BVDV), Parainfluenza Virus Type 3 (PI-3), Rabies Recombinant Nucleoprotein (rNP), Reovirus (REO), *Toxoplasma gondii*, Vesicular Stomatitis Virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

04 January 2021