

CERTIFICATE OF ANALYSIS

Bovine Respiratory Syncytial Virus (BRSV)

FITC Conjugate

Catalog No.:	CJ-F-BRSV-10ML
Lot:	P181210-001
Expiration:	01 March 2023

Quality Control Method: Direct FA using BRSV 2-well slide (catalog no. SLD-FAC-BRSV).

Specific Reaction: 2-4+ signal on the positive well and negative on the negative well, no background.

Other Comments: NA

QC Manager: Chris Cay



QC Release Date: 01 March 2019

Pattern of Fluorescence: Granular cytoplasmic fluorescence with some small syncytia.

FA Cross-Reaction Screening: The raw material has also been screened by FA and does not react with Bovine Adenovirus Type 1, 3, and 5 (BAV-1, 3, and 5), Bovine Coronavirus (BCV), Bovine Leukemia Virus (BLV), Bovine Parvovirus (BPV), Bluetongue Virus (BTV), Bovine Viral Diarrhea Virus (BVDV), Bovine Herpesvirus Type 1 (BHV-1/IBR), Parainfluenza Virus Type 3 (PI-3), Reovirus (REO), Rabies Recombinant Nucleoprotein (rNP), Vesicular Stomatitis Virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: Version #1

FOR *IN VITRO* LABORATORY USE ONLY.

12 March 2019