

**CERTIFICATE OF ANALYSIS**

# Bluetongue Virus (BTV)

FITC Conjugate MAb

|              |                   |
|--------------|-------------------|
| Catalog No.: | CJ-B-BTV-MAB-10ML |
| Lot:         | P150810-003       |
| Expiration:  | 13 August 2019    |

Quality Control Method: Direct FA using BTV 12-well slide (catalog no. SLD-IFA-BTV).

Specific Reaction: 1-4+ fluorescence with the conjugate at neat, the diluent control was negative.

Other Comments: Reacts with all U.S. and exotic serotypes. Does not react with Epizootic hemorrhagic disease virus (EHDV).

QC Manager:  Will Thurman

QC Release Date: 13 August 2015

Pattern Of Fluorescence: Multiple buckshot inclusions with dusty cytoplasmic fluorescence.

Technical Data Sheet Version: Version # 1

FOR *IN VITRO* LABORATORY USE ONLY.

H:\Quality\COAs and TDS\COA (VMRD)\Conjugate\BTV\BTV - 10 ml\CJ-F-BTV-MAB-10ML\_P150810-003\_190813.docx

14 August 2015