

CERTIFICATE OF ANALYSIS

Bluetongue Virus (BTV)


FITC Conjugate MAb

Catalog No.:	CJ-F-BTV-MAB 10ML
Lot:	P210316-001
Expiration:	07 April 2025

Quality Control Method: Direct FA using BTV 12-well slide (catalog no. SLD-IFA-BTV).

Specific Reaction: 1–4+ signal with trace background.

Other Comments: Reacts with all U.S. and exotic serotypes. Does not react with Epizootic hemorrhagic disease virus (EHDV).

QC Manager: 

David Hyatt

Pattern Of Fluorescence: Multiple buckshot inclusions with dusty cytoplasmic fluorescence.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

12 April 2021