

CERTIFICATE OF ANALYSIS

Clostridium chauvoei


FITC Conjugate

Catalog No.:	CJ-F-CCO-10ML
Lot:	P170620-002
Expiration:	24 February 2025

Quality Control Method: Direct FA using *Clostridium chauvoei* 12-well slide (catalog no. SLD-IFA-CCO).

Specific Reaction: 3–4+ signal with no background.

Other Comments: NA

QC Manager: 

David Hyatt

Pattern of Fluorescence: Individual rod-shaped bacteria. Some clumped bacteria. Flagella fluorescence alone is not indicative of a positive signal.

FA Cross-Reaction Screening: The raw material has also been screened by direct FA and does not react with *C. novyi*, *C. sordellii*, *C. septicum*, and *C. perfringens*.

Technical Data Sheet Version: 1

FOR *IN VITRO* LABORATORY USE ONLY.

H:\Master Documents\COA (VMRD)\Conjugate\C. chauvoei\C. chauvoei - 10 ml\CJ-F-CCO-10ML_P170620-002_250224.docx

11 March 2021