

CERTIFICATE OF ANALYSIS

## Canine Coronavirus (CCV)

FITC Conjugate

Catalog No.:	CJ-F-CCV-10ML
Lot:	P180323-002
Expiration:	09 April 2022

Quality Control Method: Direct FA using CCV 2-well slide (catalog no SLD-FAC-CCV).

Specific Reaction: 2 – 4+ signal on the positive well and negative on the negative well, no background.

Other Comments: NA

QC Manager: Chris Cay

QC Release Date: 10 April 2018



Pattern Of Fluorescence: Large fluorescent syncytia with some single cells in between.

FA Cross-Reaction Screening: The raw material has also been screened by FA and does not react with Canine Adenovirus Type 1 and 2 (CAV-1 and 2), Canine Coronavirus (CCV), Canine Distemper Virus (CDV), Canine Herpesvirus Type 1 (CHV-1), Canine Parainfluenza Virus Type 2 (CPI-2), Canine Parvovirus (CPV), Reovirus (REO), Rabies Recombinant Nucleoprotein (rNP), *Toxoplasma gondii*, Vesicular Stomatitis Virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: Version # 1

FOR *IN VITRO* LABORATORY USE ONLY.

18 April 2018