

CERTIFICATE OF ANALYSIS

Canine Distemper Virus (CDV)


FITC Conjugate

Catalog No.:	CJ-F-CDV-10ML
Lot:	P230208-001
Expiration:	06 March 2031

Quality Control Method: Direct FA on CDV 2-well slide (catalog no. SLD-FAC-CDV).

Specific Reaction: 2-3+ signal on the positive well and negative on the negative well. Trace background.

Other Comments: NA

QC Manager: 

David Hyatt

Pattern Of Fluorescence: Individual cells with inclusion bodies and some plasma membrane fluorescence.

FA Cross-Reaction Screening: The raw material has also been screened by FA and does not react with Canine Adenovirus Type 1 and 2 (CAV 1 and 2), Canine Coronavirus (CCV), Canine Herpesvirus Type 1 (CHV), Canine Parainfluenza Virus Type 2 (CPI-2), Canine Parvovirus (CPV), Rabies Recombinant Nucleoprotein (rNP), Reovirus (REO), *Toxoplasma gondii*, and Vesicular Stomatitis Virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: 1

FOR *IN VITRO* LABORATORY USE ONLY.

30 March 2023