

CERTIFICATE OF ANALYSIS

# Canine Distemper Virus (CDV)

FITC Conjugate MAb

Catalog No.:	CJ-F-CDV-MAB-10ML
Lot:	P170614-002
Expiration:	09 April 2025

Quality Control Method: Direct FA on CDV 2-well slide (catalog no. SLD-FAC-CDV).

Specific Reaction: 3–4+ signal on the positive well and negative on the negative well. No background.

Other Comments: NA

QC Manager:



David Hyatt

Pattern of Fluorescence: Individual cells with inclusion bodies and some plasma membrane fluorescence.

FA Cross-Reaction Screening: The raw material has also been screened by Indirect FA and does not react with Canine Adenovirus Type 1 and 2 (CAV-1 and 2), Canine Coronavirus (CCV), Canine Herpesvirus Type 1 (CHV-1), Canine Parainfluenza Virus Type 2 (CPI-2), Canine Parvovirus (CPV), Reovirus (REO), Rabies Recombinant Nucleoprotein (rNP), *Toxoplasma gondii*, Vesicular Stomatitis Virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: 1

FOR *IN VITRO* LABORATORY USE ONLY.

H:\Master Documents\COA (VMRD)\Conjugate\CDV\CDV - 10 ml mono\CJ-F-CDV-MAB-10ML\_P170614-002\_250409.docx

20 April 2021