

CERTIFICATE OF ANALYSIS

Canine Parvovirus (CPV)

FITC Conjugate MAb

Catalog No.:	CJ-F-CPV-MAB-10ML
Lot:	P221018-001
Expiration:	14 December 2030

Quality Control Method: Direct FA using CPV 2-well slide (catalog no. SLD-FAC-CPV).

Specific Reaction: 3–4+ signal on the positive well and negative on the negative well.
Trace–1+ background.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Single cells with cytoplasmic and nuclear fluorescence.

FA Cross-Reaction Screening: The raw material has also been screened by FA and does not react with canine adenovirus type 1 and 2 (CAV-1 and 2), canine coronavirus (CCV), canine distemper virus (CDV), canine herpesvirus type 1 (CHV-1), canine parainfluenza virus type 2 (CPI-2), rabies recombinant nucleoprotein (rNP), reovirus (REO), *Toxoplasma gondii*, vesicular stomatitis virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: 1

FOR IN VITRO LABORATORY USE ONLY.

12 December 2022