

CERTIFICATE OF ANALYSIS

Clostridium septicum

FITC Conjugate

Catalog No.:	CJ-F-CSE-10ML
Lot:	P170523-001
Expiration:	02 June 2021

Quality Control Method: Direct FA using *C. septicum* FA 12-well slide (catalog no. SLD-IFA-CSE).

Specific Reaction: 3-4+ signal with no background.

Other Comments: NA

QC Manager:  Will Thurman

QC Release Date: 05 June 2017

Pattern of Fluorescence: Individual rod-shaped bacteria. Some clumped bacteria. May react with large flagella.

FA Cross-Reaction Screening: The raw material has also been screened by Indirect FA and does not react with *Clostridium chauvoei*, *Clostridium novyi*, *Clostridium sordellii*, and *Clostridium perfringens*.

Technical Data Sheet Version: Version #1

FOR IN VITRO LABORATORY USE ONLY.

H:\Quality\COA (VMRD)\Conjugate\C. septicum\C. septicum - 10 ml\CJ-F-CSE-10ML_P170523-001_210602.docx

07 June 2017