

CERTIFICATE OF ANALYSIS

Clostridium septicum


FITC Conjugate

Catalog No.:	CJ-F-CSE-10ML
Lot:	P220908-002
Expiration:	04 October 2030

Quality Control Method: Direct FA using *C. septicum* FA 12-well slide (catalog no. SLD-IFA-CSE).

Specific Reaction: 3–4+ signal with no background.

Other Comments: NA

QC Manager: 

David Hyatt

Pattern Of Fluorescence: Individual rod-shaped bacteria. Some clumped bacteria. Flagella fluorescence alone is not indicative of a positive signal.

FA Cross-Reaction Screening: The raw material has also been screened by FA and has been found to react with *C. perfringens* and *C. sordellii* with trace fluorescence, but does not react with *C. chauvoei*, or *C. novyi*.

Technical Data Sheet Version: 1

FOR *IN VITRO* LABORATORY USE ONLY.

13 October 2022