

CERTIFICATE OF ANALYSIS

Clostridium sordellii

FITC Conjugate

Catalog No.:	CJ-F-CSO-10ML
Lot:	P150311-002
Expiration:	17 March 2019

Quality Control Method: Direct FA using *Clostridium sordellii* 12-well slide (catalog no. SLD-IFA-CSO).

Specific Reaction: 2-4+ fluorescence with no background.

Other Comments: NA

QC Manager: *Torrey Beeson* Torrey Beeson

QC Release Date: 17 March 2015

Pattern Of Fluorescence: Individual rod-shaped bacteria. Some clumped bacteria. Flagella fluorescence alone is not indicative of a positive signal.

FA Cross-Reaction Screening: The raw material has also been screened by direct FA and does not react with *C. novyii*, *C. septicum*, *C. perfringens*, and *C. chauvoei*.

Technical Data Sheet Version: Version # 1

FOR IN VITRO LABORATORY USE ONLY.

H:\Quality VMRD\COAs and TDS\COA (VMRD)\Conjugate\C. sordellii\C. sordellii - 10 ml\CJ-F-CSO-10ML_P150311-002_190317.docx

18 March 2015