

CERTIFICATE OF ANALYSIS

Clostridium sordellii

FITC Conjugate

Catalog No.:	CJ-F-CSO-10ML
Lot:	P180531-002
Expiration:	15 June 2022

Quality Control Method: Direct FA using *Clostridium sordellii* 12-well slide (catalog no. SLD-IFA-CSO).

Specific Reaction: 2-3+ signal with no background.

Other Comments: NA

QC Manager: Chris Cay



QC Release Date: 18 June 2018

Pattern Of Fluorescence: Individual rod-shaped bacteria. Some clumped bacteria. Flagella fluorescence alone is not indicative of a positive signal.

FA Cross-Reaction Screening: The raw material has also been screened by direct FA and does not react with *C. novyii*, *C. septicum*, *C. perfringens*, and *C. chauvoei*.

Technical Data Sheet Version: Version # 1

FOR IN VITRO LABORATORY USE ONLY.

03 July 2018