

CERTIFICATE OF ANALYSIS

Clostridium sordellii

FITC Conjugate

Catalog No.:	CJ-F-CSO-10ML
Lot:	P230320-001
Expiration:	21 March 2031

Quality Control Method: Direct FA using Clostridium spp 4-way slide (catalog no. SLD-IFA-C4).

Specific Reaction: 1–3+ signal with no background.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Individual rod-shaped bacteria. Some clumped bacteria. Flagella fluorescence alone is not indicative of a positive signal.

FA Cross-Reaction Screening: The raw material has also been screened by FA and has been found to react with *C. perfringens* with trace fluorescence, but does not react with *C. chauvoei*, *C. novyi*, or *C. septicum*.

Technical Data Sheet Version: 2

FOR IN VITRO LABORATORY USE ONLY.

28 March 2023