

CERTIFICATE OF ANALYSIS

Equine Herpesvirus Type 1 (EHV-1/ERV)

FITC Conjugate

Catalog No.:	CJ-F-ERV-10ML
Lot:	P200525-001
Expiration:	08 August 2031

Quality Control Method: Direct FA using EHV-1/ERV 2-well slide (catalog no. SLD-FAC-ERV).

Specific Reaction: 2–4+ signal on the positive well and negative on the negative well. Trace background.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Strong undifferentiated cytoplasmic fluorescence. Single cells with occasional nuclear inclusions and some small syncytia. Some intense fluorescence of dying rounded cells.

FA Cross-Reaction Screening: The raw material has also been screened by FA and does not react with Equine Arteritis Virus (EAV), Equine Infectious Anemia Virus (EIAV), reovirus (REO), Rabies Recombinant Nucleoprotein (rNP), Vesicular Stomatitis Virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

01 September 2023