

CERTIFICATE OF ANALYSIS

Anti-Feline IgG (H&L)

FITC Conjugate Affinity Purified

Catalog No.:	CJ-F-FELG-AP-10ML
Lot:	P200224-001
Expiration:	17 September 2023

Quality Control Method: Indirect FA using FIP-1 12-well slide (catalog no. SLD-IFA-FIP1), positive control (catalog no. PC-IFA-FIP1), and negative control (catalog no. NC-IFA-FIP1).

Specific Reaction: 2-4+ signal with the positive control at neat, no background. Negative with the negative control, trace background.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: The pattern of fluorescence will vary depending on what feline system was used.

FA Cross-Reaction Screening: The raw material has also been tested satisfactorily by VMRD's IFA systems for *Bartonella henselae*, Feline Calicivirus (FCV), Feline Leukemia Virus (FeLV), Feline Herpesvirus (FHV-1/FVR), Feline Infectious Peritonitis type 1 and 2 (FIP-1 and 2), Feline Panleukopenia Virus (FPLV), Reovirus (REO), Rabies Recombinant Nucleoprotein (rNP), *Toxoplasma gondii*, and Vesicular Stomatitis Virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: 1

FOR *IN VITRO* LABORATORY USE ONLY.

31 October 2019