

CERTIFICATE OF ANALYSIS

Anti-Feline IgM (heavy chain specific)

FITC Conjugate Affinity Purified

Catalog No.:	CJ-F-FELM-AP-10ML
Lot:	P190313-001
Expiration:	04 February 2023

Quality Control Method: Indirect FA using *Bartonella henselae* 12-well slide (catalog no. SLD-IFA-BH) and IgM positive control (catalog no. PC-IFA-BH-M).

Specific Reaction: 1 - 3+ signal with the positive control at neat, trace background. Negative with the diluent control, trace background.

Other Comments: NA

QC Manager: Chris Cay



QC Release Date: 04 February 2019

Pattern Of Fluorescence: The pattern of fluorescence will vary depending on what feline system was used.

FA Cross-Reaction Screening: The raw material has also been tested satisfactorily by VMRD's IFA systems for *Bartonella Henselae*, Feline Calicivirus (FCV), Feline Leukemia Virus (FeLV), Feline Herpesvirus (FHV-1/FVR), Feline Immunodeficiency Virus (FIV), Feline Infectious Peritonitis type 1 and 2 (FIP-1 and 2), Feline Panleukopenia Virus (FPLV), Reovirus (REO), and Rabies Recombinant Nucleoprotein (rNP), *Toxoplasma gondii*, Vesicular Stomatitis Virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: Version #1

FOR IN VITRO LABORATORY USE ONLY.

05 February 2019