

CERTIFICATE OF ANALYSIS

## Feline Infectious Peritonitis Virus (FIP)

FITC Conjugate

Catalog No.:	CJ-F-FIP-10ML
Lot:	P190216-001
Expiration:	04 December 2022

**Quality Control Method:** Direct FA using FIP-1 2-well slide (catalog no. SLD-FAC-FIP1) and FIP-2 2-well slide (catalog no. SLD-FAC-FIP2).

**Specific Reaction:** FIP 1 had 1 – 3+ signal and FIP 2 had 1 – 3+ signal, both with trace background, on the positive wells. Negative was negative, with trace background, on the negative wells.

**Other Comments:** NA

**QC Manager:** Chris Cay



**QC Release Date:** 04 December 2018

**Pattern Of Fluorescence:** Multiple syncytia, plaques and some individual cells with granular cytoplasmic fluorescence.

**FA Cross-Reaction Screening:** The raw material has also been screened by Indirect FA and does not react with feline calicivirus (FCV), feline leukemia virus (FeLV), feline herpesvirus (FHV-1/FVR), feline panleukopenia virus (FPLV), reovirus (REO), rabies Recombinant Nucleoprotein (rNP), vesicular stomatitis virus Indiana and New Jersey strains (VSV).

**Technical Data Sheet Version:** Version # 1

FOR IN VITRO LABORATORY USE ONLY.

07 December 2018