

CERTIFICATE OF ANALYSIS

Feline Panleukopenia Virus (FPLV)

FITC Conjugate MAb

Catalog No.:	CJ-F-FPL-MAB-10ML
Lot:	P151104-002
Expiration:	15 October 2023

Quality Control Method: Direct FA using FPLV 2-well slides (catalog no. SLD-FAC-FPL).

Specific Reaction: 2-3+ signal on the positive well with trace background, negative on the negative well with trace background.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Single cells with cytoplasmic and/or nuclear fluorescence.

FA Cross-Reaction Screening: The raw material has also been screened by FA and does not react with Feline Calicivirus (FCV), Feline Leukemia Virus (FeLV), Feline Herpesvirus (FHV-1/FVR), Feline Immunodeficiency Virus (FIV), Feline Infectious Peritonitis type 1 and 2 (FIP-1 and 2), Reovirus (REO), Rabies Recombinant Nucleoprotein (rNP), Vesicular Stomatitis Virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: Version # 1

FOR *IN VITRO* LABORATORY USE ONLY.

22 October 2019