

CERTIFICATE OF ANALYSIS

Feline Herpesvirus Type 1 (FHV-1/FVR)

FITC Conjugate

Catalog No.:	CJ-F-FVR-10ML
Lot:	P220820-001
Expiration:	19 September 2030

Quality Control Method: Direct FA using FHV-1/FVR 2-well slide (catalog no. SLD-FAC-FVR).

Specific Reaction: 3–4+ signal on the positive well and negative on the negative well. No background.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Individual rounded cells with cytoplasmic and nuclear fluorescence. Some small syncytia.

FA Cross-Reaction Screening: The raw material has also been screened by FA and does not react with feline calicivirus (FCV), feline leukemia virus (FeLV), feline infectious peritonitis type 1 and 2 (FIP-1 and 2), feline panleukopenia virus (FPLV), feline immunodeficiency virus (FIV), rabies Recombinant Nucleoprotein (rNP), reovirus (REO), *Toxoplasma gondii*, vesicular stomatitis virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: 2

FOR IN VITRO LABORATORY USE ONLY.

06 October 2022