

**CERTIFICATE OF ANALYSIS**

# Bovine Herpesvirus Type 1 (BHV-1/IBR)

FITC Conjugate

Catalog No.:	CJ-F-IBR-10ML
Lot:	P181125-001
Expiration:	27 September 2030

**Quality Control Method:** Direct FA using BHV-1/IBR 2-well slide (catalog no. SLD-FAC-IBR).

**Specific Reaction:** 2–4+ signal on the positive well with trace–1+ background. Negative with trace–1+ background on the negative well.

**Other Comments:** NA

**QC Manager:**



David Hyatt

**Pattern Of Fluorescence:** Primarily undifferentiated cytoplasmic with some nuclear fluorescence, especially in rounded cells and degenerating cells in plaques with acellular centers.

**FA Cross-Reaction Screening:** The raw material has also been screened by FA and does not react with Bovine Adenovirus Type 1, 3, and 5 (BAV-1, 3, and 5), Bovine Coronavirus (BCV), Bovine Leukemia Virus (BLV), Bovine Parvovirus (BPV), Bovine Respiratory Syncytial Virus (BRSV), Bluetongue Virus (BTV), Bovine Viral Diarrhea Virus (BVDV), Parainfluenza Virus Type 3 (PI-3), Reovirus (REO), Rabies Recombinant Nucleoprotein (rNP), Vesicular Stomatitis Virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: 1

FOR *IN VITRO* LABORATORY USE ONLY.

12 October 2022