

CERTIFICATE OF ANALYSIS

Bovine Herpesvirus Type 1 (BHV-1/IBR)


FITC Conjugate

Catalog No.:	CJ-F-IBR-10ML
Lot:	P221116-001
Expiration:	01 December 2030

Quality Control Method: Direct FA using BHV-1/IBR 2-well slide (catalog no. SLD-FAC-IBR).

Specific Reaction: 2-4+ signal on the positive well and negative on the negative well.
Trace-1+ background.

Other Comments: NA

QC Manager: 

David Hyatt

Pattern Of Fluorescence: Primarily undifferentiated cytoplasmic with some nuclear fluorescence, especially in rounded cells and degenerating cells in plaques with acellular centers.

FA Cross-Reaction Screening: The raw material has also been screened by FA and does not react with Bovine Adenovirus Type 1, 3, and 5 (BAV-1, 3, and 5), Bovine Coronavirus (BCV), Bovine Leukemia Virus (BLV), Bovine Parvovirus (BPV), Bovine Respiratory Syncytial Virus (BRSV), Bluetongue Virus (BTV), Bovine Viral Diarrhea Virus (BVDV), Parainfluenza Virus Type 3 (PI-3), Rabies Recombinant Nucleoprotein (rNP), Reovirus (REO), *Toxoplasma gondii*, Vesicular Stomatitis Virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: 1

FOR *IN VITRO* LABORATORY USE ONLY.

16 December 2022