

CERTIFICATE OF ANALYSIS

Porcine Circovirus Type 1 & 2 (PCV-1&2)

FITC Conjugate

Catalog No.:	CJ-F-PCV1&2-10ML
Lot:	P171214-003
Expiration:	18 January 2022

Quality Control Method: Direct FA using PCV-1 2-well slide (catalog no. SLD-FAC-PCV1) and PCV-2 2-well slide (catalog no. SLD-FAC-PCV2).

Specific Reaction: 2 – 4+ signal on the positive well and negative on the negative well, no background.

Other Comments: NA

QC Manager:  Will Thurman

QC Release Date: 22 January 2018

Pattern Of Fluorescence: Intense whole nuclear and variable sized but discrete cytoplasmic inclusions.

FA Cross-Reaction Screening: The raw material has also been screened by FA and does not react with Porcine Adenovirus (PAV), Porcine Hemagglutinating Encephalomyelitis Virus (PHEV), Porcine Parvovirus (PPV), Porcine Reproductive and Respiratory Syndrome Virus (PRRSV) North American and European strains, Transmissible Gastroenteritis Virus (TGEV), Reovirus (REO), Rabies Recombinant Nucleoprotein (rNP), and Vesicular Stomatitis Virus (VSV) Indiana and New Jersey strains.

Technical Data Sheet Version: Version # 2

FOR IN VITRO LABORATORY USE ONLY.

H:\Master Documents\COA (VMRD)\Conjugate\PCV1&2\PCV1&2 10 ML\CJ-F-PCV12-10ML_P171214-003_220118.docx

15 February 2018