

CERTIFICATE OF ANALYSIS

Porcine Circovirus Type 2 (PCV-2)

FITC Conjugate

Catalog No.:	CJ-F-PCV2-10ML
Lot:	P221217-001
Expiration:	19 January 2031

Quality Control Method: Direct FA using PCV-2 2-well slide (catalog no. SLD-FAC-PCV2).

Specific Reaction: 3–4+ signal on the positive well and negative on the negative well. No background.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Intense whole nuclear and variable sized but discrete cytoplasmic inclusions.

FA Cross-Reaction Screening: The raw material has also been screened by FA and has been found to react with Porcine Adenovirus (PAV) at trace–1+, but does not react with Porcine Circovirus Type 1 (PCV-1), Porcine Hemagglutinating Encephalomyelitis Virus (PHEV), Porcine Parvovirus (PPV), Porcine Reproductive and Respiratory Syndrome Virus (PRRSV) North American and European strains, Transmissible Gastroenteritis Virus (TGEV), Rabies Recombinant Nucleoprotein (rNP), Reovirus (REO), *Toxoplasma gondii*, Vesicular Stomatitis Virus (VSV) Indiana and New Jersey strains.

Technical Data Sheet Version: 2

FOR IN VITRO LABORATORY USE ONLY.

01 February 2023