

CERTIFICATE OF ANALYSIS

Parainfluenza Virus Type 3 (PI-3)

FITC Conjugate

Catalog No.:	CJ-F-PI3-10ML
Lot:	P200110-001
Expiration:	13 February 2024

Quality Control Method: Direct FA using PI-3 2-well slide (catalog no. SLD-FAC-PI3).

Specific Reaction: 2-4+ signal on the positive well and negative on the negative well, no background.

Other Comments: NA

QC Manager: 

David Hyatt

Pattern Of Fluorescence: Small, bright spherical cytoplasmic inclusions in syncytia and single cells.

FA Cross-Reaction Screening: The raw material has also been screened by FA and does not react with Bovine Adenovirus Type 1, 3, and 5 (BAV-1, 3, and 5), Bovine Coronavirus (BCV), Bovine Herpesvirus type 1 (BHV-1/IBR), Bovine Leukemia Virus (BLV), Bovine Parvovirus (BPV), Bovine Respiratory Syncytial Virus (BRSV), Bluetongue Virus (BTV), Bovine Viral Diarrhea Virus (BVDV), Rabies Recombinant Nucleoprotein (rNP), Reovirus (REO), *Toxoplasma gondii*, Vesicular Stomatitis Virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: 1

FOR IN VITRO LABORATORY USE ONLY.

24 February 2020