

CERTIFICATE OF ANALYSIS

Reovirus (REO)

FITC Conjugate

Catalog No.:	CJ-F-REO-10ML
Lot:	P200130-001
Expiration:	25 November 2023

Quality Control Method: Direct FA using REO 2-well slide (catalog no. SLD-FAC-REO).

2-4+ signal with the positive control at neat, no background. Negative with Specific Reaction:

the negative control, no background.

Other Comments:

QC Manager:

David Hyatt

Pattern Of Fluorescence: Individual cells or plaques with diffuse cytoplasmic fluorescence and

characteristic bright cytoplasmic inclusions ranging in size from medium to

large.

Screening:

FA Cross-Reaction The raw material has also been screened by FA and does not react with Bovine Adenovirus Type 1, 3, and 5 (BAV-1, 3, and 5), Bovine Coronavirus

(BCV), Bovine Leukemia Virus (BLV), Bovine Parvovirus (BPV), Bovine Respiratory Syncytial Virus (BRSV), Bluetongue Virus (BTV), Bovine Viral

Diarrhea Virus (BVDV), Bovine Herpesvirus type 1 (BHV-1/IBR),

Parainfluenza Virus Type 3 (PI-3), Rabies Recombinant Nucleoprotein (rNP),

Vesicular Stomatitis Virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: 1

FOR IN VITRO LABORATORY USE ONLY.