

CERTIFICATE OF ANALYSIS

Reovirus (REO)

FITC Conjugate

Catalog No.:	CJ-F-REO-1ML
Lot:	P180219-003
Expiration:	21 February 2022

Quality Control Method: Direct FA using REO 2-well slide (catalog no. SLD-FAC-REO).

Specific Reaction: 3 – 4+ signal on the positive well and negative on the negative well, no background.

Other Comments: NA

QC Manager:  Will Thurman

QC Release Date: 21 February 2018

Pattern Of Fluorescence: Individual cells or plaques with diffuse cytoplasmic fluorescence and characteristic bright cytoplasmic inclusions ranging in size from medium to large.

FA Cross-Reaction Screening: The raw material has also been screened by FA and does not react with Bovine Adenovirus Type 1, 3, and 5 (BAV-1, 3, and 5), Bovine Coronavirus (BCV), Bovine Leukemia Virus (BLV), Bovine Parvovirus (BPV), Bovine Respiratory Syncytial Virus (BRSV), Bluetongue Virus (BTV), Bovine Viral Diarrhea Virus (BVDV), Bovine Herpesvirus type 1 (BHV-1/IBR), Parainfluenza Virus Type 3 (PI-3), Rabies Recombinant Nucleoprotein (rNP), Vesicular Stomatitis Virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: Version # 1

FOR IN VITRO LABORATORY USE ONLY.

H:\Master Documents\COA (VMRD)\Conjugate\Reovirus\Reovirus - 1 ml\CJ-F-REO-1ML_P180219-003_2202021.docx

06 March 2018