

CERTIFICATE OF ANALYSIS

Vesicular Stomatitis Virus (VSV)

FITC Conjugate

Catalog No.:	CJ-F-VSV-10ML
Lot:	P160916-001
Expiration:	12 October 2020

Quality Control Method: Direct FA on VSV 12-well Indiana and New Jersey strains in house slides.

Specific Reaction: 4+ signal and no background with the New Jersey Strain. 3-4+ signal and no background with the Indiana Strain.

Other Comments: NA

QC Manager:  Will Thurman

QC Release Date: 13 October 2016

Pattern of Fluorescence: Plaques of cells exhibiting fine grained ground glass cytoplasmic fluorescence with some larger cytoplasmic inclusions.

FA Cross-Reaction Screening: The raw material has also been screened by FA and does not react with Porcine Adenovirus (PAV), Porcine Cytomegalovirus (PCMV), Porcine Circovirus Type 1 and 2 (PCV-1 and 2), Porcine Hemagglutinating Encephalomyelitis Virus (PHEV), Porcine Parvovirus (PPV), Porcine Reproductive and Respiratory Syndrome Virus (PRRSV) North American and European strains, Transmissible Gastroenteritis Virus (TGEV), Reovirus (REO), Rabies Recombinant Nucleoprotein (rNP).

Technical Data Sheet Version: Version # 1

FOR IN VITRO LABORATORY USE ONLY.

H:\Quality\COA (VMRD)\Conjugate\VSV\VSV - 10 ml\CJ-F-VSV-10ML_P160916-001_201012.docx

24 October 2016