

CERTIFICATE OF ANALYSIS

Vesicular Stomatitis Virus (VSV)

FITC Conjugate

Catalog No.:	CJ-F-VSV-10ML
Lot:	P190706-001
Expiration:	07 May 2023

Quality Control Method: Direct FA on VSV 12-well Indiana and 12-well New Jersey strains in house slides.

Specific Reaction: 2-4+ signal with no background on Indiana strain. 3-4+ signal with no background on the New Jersey strain.

Other Comments: NA

QC Manager: Chris Cay



QC Release Date: 08 May 2019

Pattern Of Fluorescence: Plaques of cells exhibiting fine grained ground glass cytoplasmic fluorescence with some larger cytoplasmic inclusions.

FA Cross-Reaction Screening: The raw material has also been screened by FA and does not react with Porcine Adenovirus (PAV), Porcine Cytomegalovirus (PCMV), Porcine Circovirus Type 1 and 2 (PCV-1 and 2), Porcine Hemagglutinating Encephalomyelitis Virus (PHEV), Porcine Parvovirus (PPV), Porcine Reproductive and Respiratory Syndrome Virus (PRRSV) North American and European strains, Transmissible Gastroenteritis Virus (TGEV), Reovirus (REO), Rabies Recombinant Nucleoprotein (rNP).

Technical Data Sheet Version: Version #1

FOR *IN VITRO* LABORATORY USE ONLY.

3 June 2019