

CERTIFICATE OF ANALYSIS

Vesicular Stomatitis Virus (VSV)

FITC Conjugate

Catalog No.:	CJ-F-VSV-10ML
Lot:	P201021-001
Expiration:	20 November 2024

Quality Control Method: Direct FA on VSV 12-well Indiana and New Jersey strains in house slides.

Specific Reaction: 3–4+ signal with no background on both the Indiana and New Jersey strains.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Plaques of cells exhibiting fine grained ground glass cytoplasmic fluorescence with some larger cytoplasmic inclusions.

FA Cross-Reaction Screening: The raw material has also been screened by FA and has been found to react with Porcine Adenovirus (PAV) with trace signal, but does not react with Bovine Adenovirus Type 1, 3, and 5 (BAV-1, 3, and 5), Bovine Coronavirus (BCV), Bovine Herpesvirus type 1 (BHV-1/IBR), Bovine Leukemia Virus (BLV), Bovine Parvovirus (BPV), Bovine Respiratory Syncytial Virus (BRSV), Bluetongue Virus (BTV), Bovine Viral Diarrhea Virus (BVDV), Equine Arteritis Virus (EAV), Equine Herpesvirus Type 1 (EHV-1/ERV), Porcine Adenovirus (PAV), Porcine Circovirus Type 1 and 2 (PCV-1 and 2), Porcine Hemagglutinating Encephalomyelitis Virus (PHEV), Porcine Parvovirus (PPV), Porcine Reproductive and Respiratory Syndrome Virus (PRRSV) North American and European strains, Transmissible Gastroenteritis Virus (TGEV), Parainfluenza Virus Type 3 (PI-3), Rabies Recombinant Nucleoprotein (rNP), Reovirus (REO), and *Toxoplasma gondii*.

Technical Data Sheet Version: 1

FOR *IN VITRO* LABORATORY USE ONLY.

14 December 2020