

CERTIFICATE OF ANALYSIS

Clostridium novyi

FITC Conjugate

Catalog No.:	CJ-F-CNO-10ML
Lot:	P160606-001
Expiration:	08 June 2020

Quality Control Method: Direct FA using *C. novyi* 12-well slide (catalog no. SLD-IFA-CNO).

Specific Reaction: 3-4+ signal with no background.

Other Comments: NA

QC Manager:  Will Thurman

QC Release Date: 08 June 2016

Pattern Of Fluorescence: Individual rod-shaped bacteria. Some clumped bacteria. May react with large flagella.

FA Cross-Reaction Screening: The raw material has also been screened by FA and does not react with *C. chauvoei*, *C. perfringens*, *C. septicum* or *C. sordellii*.

Technical Data Sheet Version: Version # 1

FOR IN VITRO LABORATORY USE ONLY.

H:\Quality\COA (VMRD)\Conjugate\C. novyi\C. novyi - 10 ml\CJ-F-CNO-10ML_P160606-001_200608.docx

09 June 2016