

CERTIFICATE OF ANALYSIS

FA Mounting Fluid

Catalog No.:	FAMF-10ML
Lot:	P181205-001
Expiration:	22 January 2023

Quality Control Method: Direct FA using BHV-1/IBR 2-well slides (catalog no. SLD-FAC-IBR) and FITC conjugate (catalog number: CJ-F-IBR-1ML or 10ML).

Specific Reaction: There is no difference between the reference and the QC lot of mounting fluid.

pH: 9.46

Other Comments: NA

QC Manager: Chris Cay



QC Release Date: 22 January 2019

Technical Data Sheet Version: Version # 1

FOR *IN VITRO* LABORATORY USE ONLY.

25 January 2019