

CERTIFICATE OF ANALYSIS

4X FA Powdered Rinse Buffer

Catalog No.:	FARB-4X
Lot:	P160502-001
Expiration:	25 May 2020

Quality Control Method: Reconstitute in 1 liter of deionized water to make 4X concentrate. Using pH meter, determine the pH of the solution. Solution should be between pH 9.0 and pH 9.5. Dilute to 1X. Direct FA using FIP conjugate (catalog no. CJ-F-FIP-1ML or 10ML) and FIP1 2-well control slide (catalog no. SLD-FAC-FIP1).

Specific Reaction: Reconstituted 4X and 1X rinse buffer was clear. QC rinse buffer produced results identical to reference rinse buffer.

pH: 9.06

Other Comments: NA

QC Manager:  Will Thurman

QC Release Date: 25 May 2016

Technical Data Sheet Version: Version #1

FOR IN VITRO LABORATORY USE ONLY.

H:\Quality\COA (VMRD)\Buffers\Rinse Buffer\FARB-4X_P160502-001_200525.docx

31 May 2016