

**VMRD**

PO Box 502, Pullman, WA 99163 USA

Telephone: 509-334-5815
Fax: 509-332-5356
E-mail: vmrd@vmrd.com
Web site: <http://www.vmrd.com>

Certificate of Analysis

BOVINE RESPIRATORY SYNCYTIAL VIRUS (BRSV)

Negative Control Serum

CATALOG NO.: 211-N-BRSV
re-order as NC-IFA-BRSV

SIZE: 1 ml

LOT: P081114-001

EXPIRATION: 04 January 2032

VIRUS: Bovine Respiratory Syncytial Virus (BRSV)

DESCRIPTION: Bovine serum diluted in PBS, 1% BSA, 0.09% sodium azide.

QUALITY CONTROL METHOD: Indirect FA using VMRD, Inc. BRSV 12-well (catalog no. SLD-IFA-BRSV), and Anti-Bovine IgG_{1,2} FITC Conjugate (catalog no. CJ-F-BOVG-AP-10ML or 1ML).

Specific Reaction: Negative with trace background.

Other Comments: NA

PATTERN OF FLUORESCENCE: None with our system.

INTENDED USE: As negative control serum in detection of antibody to Bovine Respiratory Syncytial Virus (BRSV) by indirect FA technique. This serum should be used undiluted to demonstrate non-specific fluorescence, if any.

STABILITY: Store at 2-7°C. DO NOT FREEZE.

FOR *IN VITRO* LABORATORY USE ONLY.

WARRANTY: VMRD, Inc. warrants that this product is as described in the quantity and contents stated on the label at the time of delivery to the customer. NO OTHER WARRANTIES, EXPRESS OR IMPLIED, ARE MADE BEYOND THE LABEL DESCRIPTION, INCLUDING WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR USE. Remedy is limited to replacement of the product or refund of the purchase price. VMRD, Inc. is not liable for property damage, personal injury, or economic loss caused by the product. The information listed in this information sheet is provided for reference only, and should not be substituted for the user's own incoming material quality control.

[https://vmrd1.sharepoint.com/quality/Master Documens/COA \(VMRD\)/Negative Controls/BRSV/NC-IFA-BRSV_P081114-001_320104.docx](https://vmrd1.sharepoint.com/quality/Master%20Documens/COA%20(VMRD)/Negative%20Controls/BRSV/NC-IFA-BRSV_P081114-001_320104.docx)
10 January 2024