

CERTIFICATE OF ANALYSIS

Bovine Viral Diarrhea Virus (BVDV)

Antiserum

Catalog No.:	PAB-BVD
Lot:	P200311-001
Expiration:	14 August 2024

Quality Control Method: Indirect FA using BVDV 12-well slide (catalog no. SLD-IFA-BVD) and anti-caprine IgG FITC conjugate (catalog no. CJ-F-CAPG-10ML).

Specific Reaction: 2–4+ signal with the antisera at 1/400 dilution, trace background. Endpoint titer is 1/1,600.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Individual cells with smooth, undifferentiated and/or “ground glass” cytoplasmic fluorescence.

FA Cross-Reaction Screening: The Antiserum has also been screened by Indirect FA and does not react with Bovine Adenovirus Type 1, 3, and 5 (BAV-1, 3, and 5), Bovine Coronavirus (BCV), Bovine Herpesvirus type 1 (BHV-1/IBR), Bovine Leukemia Virus (BLV), Bovine Parvovirus (BPV), Bovine Respiratory Syncytial Virus (BRSV), Bluetongue Virus (BTV), Parainfluenza Virus Type 3 (PI-3), Rabies Recombinant Nucleoprotein (rNP), Reovirus (REO), *Toxoplasma gondii*, Vesicular Stomatitis Virus Indiana and New Jersey strains (VSV).

Technical Data Sheet Version: 1

FOR IN VITRO LABORATORY USE ONLY.

28 August 2020