

CERTIFICATE OF ANALYSIS

Bovine Respiratory Syncytial Virus (BRSV)

FA Positive Control

Catalog No.:	PC-IFA-BRSV
Lot:	P170310-002
Expiration:	17 December 2024

Quality Control Method: IFA using VMRD, Inc. BRSV 12-well slide (catalog no. SLD-IFA-BRSV), BRSV Negative Control (catalog no. NC-IFA-BRSV), and Anti-Bovine IgG1,2 FITC Conjugate (catalog no. CJ-F-BOVG-AP-10ML).

Specific Reaction: 2-4+ signal with the positive control at neat, trace background. Endpoint reached in 3 two-fold dilutions.

Other Comments: NA

QC Manager: 

David Hyatt

Pattern of Fluorescence: Granular cytoplasmic fluorescence with some small syncytia.

Technical Data Sheet Version: 1

FOR *IN VITRO* LABORATORY USE ONLY.

H:\Master Documents\COA (VMRD)\Positive Controls\BRSV\PC-IFA-BRSV_P170310-002_241217.docx

29 December 2020