

CERTIFICATE OF ANALYSIS

Brucella canis

FA Positive Control

Catalog No.:	PC-IFA-CB
Lot:	P190705-001
Expiration:	04 June 2023

Quality Control Method: IFA using *Brucella canis* 12-well slide (catalog no. SLD-IFA-CB), negative control (catalog no. NC-IFA-CB), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-1ML or 10ML).

Specific Reaction: 2-3+ signal with the control at neat, no background. Endpoint is trace - 1+ at 1/8 dilution.

Other Comments: NA

QC Manager: David Hyatt



QC Release Date: 04 June 2019

Pattern Of Fluorescence: Brightly staining short bacilli against uniform background of ovine red blood cells.

Technical Data Sheet Version: Version #1

FOR *IN VITRO* LABORATORY USE ONLY.

14 June 2019