

CERTIFICATE OF ANALYSIS

*Brucella canis*


FA Positive Control

Catalog No.:	PC-IFA-CB
Lot:	P221023-001
Expiration:	23 November 2030

**Quality Control Method:** Indirect FA using *Brucella canis* 12-well slide (catalog no. SLD-IFA-CB), negative control (catalog no. NC-IFA-CB), and anti-canine IgG affinity purified FITC conjugate (catalog no. CJ-F-CANG-AP-10ML).

**Specific Reaction:** 3–4+ signal with the control at neat. No background. Endpoint signal was reached at 1/4 dilution.

**Other Comments:** NA

**QC Manager:** 

David Hyatt

**Pattern Of Fluorescence:** Brightly staining short bacilli against uniform background of ovine red blood cells.

Technical Data Sheet Version: 1