

CERTIFICATE OF ANALYSIS

# Canine Distemper Virus (CDV) IgG

FA Positive Control

Catalog No.:	PC-IFA-CDV-G
Lot:	P211015-001
Expiration:	11 November 2025

**Quality Control Method:** Indirect FA on CDV 12-well slide (catalog no. SLD-IFA-CDV), and anti-canine IgG FITC conjugates (catalog no. CJ-F-CANG-10ML).

**Specific Reaction:** 2–4+ signal with the control at neat with trace–1+ background. Endpoint signal was reached at greater than 1/16 dilution.

**Other Comments:** NA

**QC Manager:**



David Hyatt

**Pattern Of Fluorescence:** Individual cells with inclusion bodies and some plasma membrane fluorescence. High titer test sera will fluoresce brighter than this positive control serum.

Technical Data Sheet Version: 1

FOR *IN VITRO* LABORATORY USE ONLY.

18 November 2021