

**CERTIFICATE OF ANALYSIS**

# Canine Distemper Virus (CDV) IgM

FA Positive Control

Catalog No.:	PC-IFA-CDV-M
Lot:	P180716-003
Expiration:	25 July 2022

**Quality Control Method:** Indirect FA on CDV 12-well slide (catalog no. SLD-IFA-CDV), and anti-canine IgM affinity purified conjugates (catalog no. CJ-F-CANM-AP-1ML or 10ML).

**Specific Reaction:** 2-4+ signal with the control at neat, no background. Endpoint is trace to 1+ at 1/4 dilution.

**Other Comments:** NA

**QC Manager:** Chris Cay



**QC Release Date:** 25 July 2018

**Pattern Of Fluorescence:** Individual cells with inclusion bodies and some plasma membrane fluorescence. High titer test sera will fluoresce brighter than this positive control serum.

**Technical Data Sheet Version:** Version # 1

FOR *IN VITRO* LABORATORY USE ONLY.

02 August 2018