

CERTIFICATE OF ANALYSIS

# Canine Distemper Virus (CDV) IgM

FA Positive Control

Catalog No.:	PC-IFA-CDV-M
Lot:	P201007-001
Expiration:	23 October 2024

**Quality Control Method:** Indirect FA on CDV 12-well slide (catalog no. SLD-IFA-CDV), and anti-canine IgM affinity purified conjugates (catalog no. CJ-F-CANM-AP-10ML).

**Specific Reaction:** 3–4+ signal with the control at neat, trace background. Endpoint signal was reached at 1/8 dilution.

**Other Comments:** NA

**QC Manager:**



David Hyatt

**Pattern Of Fluorescence:** Individual cells with inclusion bodies and some plasma membrane fluorescence. High titer test sera will fluoresce brighter than this positive control serum.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

05 November 2020