

**CERTIFICATE OF ANALYSIS**

# Canine Herpesvirus Type 1 (CHV-1)


FA Positive Control

Catalog No.:	PC-IFA-CHV
Lot:	P200312-001
Expiration:	18 March 2032

**Quality Control Method:** Indirect FA using CHV-1 12-well slide (catalog no. SLD-IFA-CHV) and anti-canine IgG affinity purified FITC conjugate (catalog no. CJ-F-CANG-AP-10ML).

**Specific Reaction:** 2–3+ signal with the control at neat. Trace background. Endpoint signal was reached at 1/8 dilution.

**Other Comments:** NA

**QC Manager:** 

David Hyatt

**Pattern Of Fluorescence:** Cytoplasmic and nuclear fluorescence in rounded cells with occasional small syncytia.

Technical Data Sheet Version: 1

FOR *IN VITRO* LABORATORY USE ONLY.

20 March 2024